

TWD Scientific, LLC 10555 86th Avenue Pleasant Prairie, WI 53158

888.323.3585 TEL 262.605.3262 FAX

www.twdtradewinds.com

PART #:	LOT #:
DF8988P-S	202304015
DESCRIPTION:	
DISPOSABLE REINFORCED NYLON STERILE WHITE FORCEPS WITH BLUNT TIP. FORCEPS ARE INDIVIDUALLY PEEL-PACKED. *INTENDED FOR SINGLE USE ONLY*	
SPECIFICATIONS:	MATERIAL:
LENGTH: 122.5mm	
WIDTH: 22.5mm	REINFORCED NYLON
TIP OPENING: 15mm	
EXPIRATION DATE	STERILIZATION DATE
(MONTH/DAY/YEAR): 04/23/2028	(MONTH/DAY/YEAR): 04/23/2023

CERTIFICATE OF STERILIZATION

The validation and routine control of the sterilization process are carried out in accordance with the following standards:

ISO 11135:

Sterilization of health care products – Ethylene oxide – Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices

Provided appropriately stored, products remain sterile for at least 5 years from the sterilization date.

CERTIFICATE OF CONFORMITY

We hereby certify that product mention above meets the specifications stated.

Lot #: 20230401S

Expiration (Month/Day/Year): 04/23/2028