



TWD Scientific, LLC
10555 86th Avenue
Pleasant Prairie, WI 53158

888.323.3585 TEL | 262.605.3262 FAX

www.twdtradewinds.com

PART #:	REF #:	DESCRIPTION:	LOT #:
DF8988P-S-ENDO	84012601	ENDOTOXIN-FREE DISPOSABLE REINFORCED NYLON STERILE WHITE FORCEPS WITH BLUNT TIP. FORCEPS ARE INDIVIDUALLY PEEL-PACKED.	357150
STERILIZATION DATE: 03/2024		EXPIRATION DATE: 02/2029	

CERTIFICATE OF STERILIZATION

The validation and routine control of the sterilization process are carried out in accordance with the following standards:

EN ISO 11135: 2014	Sterilization of health care products – Ethylene oxide – Part 1: Requirements for development, validation and routine control of a sterilizations process for medical devices
EN 556-1: 2001	Sterilization of medical devices – Requirements for medical devices to be designated “Sterile” – Part 1: Requirements for terminally sterilized medical devices
EN ISO 10993-7: 2008	Biological evaluation of medical devices – Part 7: Ethylene oxide sterilization residuals
Routine control:	The criteria for product release from sterilization - Conformity with the specified physical parameters for the sterilization cycle - Inactivation of biological indicators

LAL testing is carried out in accordance with the following standards:

GOST ISO 10993-11-2021	Biological evaluation of medical devices Part 11: Tests for systemic toxicity
------------------------	---

Provided appropriately stored, products remain sterile for at least 5 years from the sterilization date.

CERTIFICATE OF CONFORMITY

Has been developed, manufactured, inspected and sterilized in accordance with the requirements of:

ISO 13485:2016: Quality systems – Medical devices.

EN ISO 13485:2016: Quality systems – Medical devices.